RayBio® Human Phospho-mTOR (S2448) and Total mTOR ELISA Kit

For Measuring Phosphorylated mTOR (S2448) and Total mTOR in Human Cell Lysates

User Manual
(Revised May 16th, 2014)

RayBio® Phospho-mTOR (S2448) and Total mTOR ELISA Kit Protocol

(Cat#: PEL-mTOR-S2448-T)

RayBiotech, Inc.

ISO 13485 & GLP Certified

Tel: (Toll Free) 1-888-494-8555 or 770-729-2992; Fax: 770-206-2393;
Web: www.raybiotech.com Email: info@raybiotech.com

RayBiotech, Inc.

RayBio® Phospho-mTOR (S2448) and Total mTOR ELISA Kit Protocol

TABLE OF CONTENTS

I. Introduction……..……2
II. Material Provided……..……..……3
III. Storage..……3
IV. Additional Materials Required……4
V. Sample Preparation……...4
VI. Reagent Preparation……..5
VII. Assay Procedure……....7
VIII. Assay Procedure Summary……...8
IX. Typical Data……..9
i. Positive Control……9
ii. EGF Stimulation of A431 Cell Lines……..10
X. Troubleshooting Guide……...……..11

I. INTRODUCTION

RayBio® Phospho- mTOR (S2448) and Total mTOR ELISA kit is a very rapid, convenient and sensitive assay kit that can monitor the activation or function of important biological pathways in human cell lysates. By determining phosphorylated mTOR protein in your experimental model system, you can verify pathway activation in your cell lysates. You can simultaneously measure numerous different cell lysates without spending excess time and effort in performing a Western Blotting analysis.

This Sandwich ELISA kit is an in vitro enzyme-linked immunosorbent assay for the measurement of human phospho- mTOR and total mTOR. An anti-pan mTOR antibody has been coated onto a 96-well plate. Samples are pipetted into the wells and mTOR present in a sample is bound to the wells by the immobilized antibody. The wells are washed and rabbit anti-mTOR (S2448) antibody is used to detect phosphorylated mTOR or biotinylated anti-mTOR antibody is used to detect pan mTOR. After washing away unbound antibody, HRP-conjugated anti-rabbit IgG or HRP-conjugated Streptavidin is pipetted to the wells. The wells are again washed, a TMB substrate solution is added to the wells and color develops in proportion to the amount of mTOR (S2448) or pan mTOR bound. The Stop Solution changes the color from blue to yellow, and the intensity of the color is measured at 450 nm.

II. MATERIAL PROVIDED

1. mTOR Microplate (Item A): 96 wells (12 strips x 8 wells) coated with anti-pan mTOR antibody.
2. Wash Buffer Concentrate (20x) (Item B): 25 ml of 20x concentrated solution.
3. Assay Diluent (Item E2): 15 ml of 5x concentrated buffer. For diluting cell lysate sample, detection antibody (Item C) and HRP-conjugated IgG Concentrate (Item D).
4. Detection Antibody mTOR (S2448) (Item C-1): 1 vial of rabbit anti-mTOR (S2448) (1 vial is enough to assay half microplate).
5. Detection Antibody mTOR (Item C-2): 1 vial of biotinylated anti-mTOR (1 vial is enough to assay half microplate).
6. HRP-conjugated Anti-rabbit IgG (Item D-1), 25 µl of 1000x concentrated HRP-conjugated anti-rabbit IgG.
7. HRP-conjugated Streptavidin (Item G), 25 µl of 300x concentrated HRP-conjugated Streptavidin.
8. TMB One-Step Substrate Reagent (Item H): 12 ml of 3,3’,5,5’-tetramethylbenzidine (TMB) in buffered solution.
9. Stop Solution (Item I): 8 ml of 0.2 M sulfuric acid.
10. Cell Lysate Buffer (Item J): 5 ml 2x cell lysis buffer (not including protease and phosphatase inhibitors).
11. Positive Control A431S001-1 (Item K): 1 vial of lyophilized powder from A431 cell lysate.

III. STORAGE

Upon receipt, the kit should be stored at –20°C. Please use within 6 months from the date of shipment. After initial use, Wash Buffer Concentrate (Item B), Assay Diluent (Item E2), TMB One-Step Substrate Reagent (Item H), Stop Solution (Item I) and Cell Lysate Buffer (Item J) should be stored at 4°C to avoid repeated freeze-thaw cycles. Return unused wells to the pouch containing desiccant pack, reseal along entire edge and store at –20°C. Item D and G store at 2-8°C for up to one month (store at -20°C for up to 6 months, avoid repeated freeze-thaw cycles). Reconstituted Positive Control (Item K) should be stored at -70°C.

IV. ADDITIONAL MATERIALS REQUIRED

1. Microplate reader capable of measuring absorbance at 450 nm.
2. Protease and Phosphatase inhibitors.
3. Shaker.
4. Precision pipettes to deliver 2 μl to 1 ml volumes.
5. Adjustable 1-25 ml pipettes for reagent preparation.
6. 100 ml and 1 liter graduated cylinders.
7. Distilled or deionized water.
8. Tubes to prepare sample dilutions.

V. SAMPLE PREPARATION

Cell lysates - Rinse cells with PBS, making sure to remove any remaining PBS before adding the lysis buffer. Solubilize cells at 4 x 10^7 cells/ml in 1x Lysis Buffer (we recommend adding protease and phosphatase inhibitors to lysis buffer prior to sample preparation). Pipette up and down to resuspend and incubate the lysates with shaking at 2 - 8°C for 30 minutes. Microcentrifuge at 13,000 rpm for 10 minutes at 2 - 8°C, and transfer the supernates into a clean test tube. Lysates should be used immediately or aliquoted and stored at –70°C. Avoid repeated freeze-thaw cycles. Thawed lysates should be kept on ice prior to use.

For the initial experiment, we recommend a serial dilution, such as 5-fold to 50-fold, for your cell lysates with Assay Diluent (Item E2) before use.

Note: The fold dilution of sample used depends on the abundance of phosphorylated proteins and should be determined empirically. More of the sample can be used if signals are too weak. If signals are too strong, the sample can be diluted further.

Cell lysate buffer should be diluted 2-fold with deionized or distilled water before use (recommend to add protease and phosphatase inhibitors).

VI. REAGENT PREPARATION

1. Bring all reagents and samples to room temperature (18 - 25°C) before use.

2. Assay Diluent (Item E2) should be diluted 5-fold with deionized or distilled water before use.

3. Preparation of Positive Control: Briefly spin the Positive Control vial of Item K. Add 400 µl 1x Assay Diluent (Item E2) into Item K to prepare Positive Control (P-1) solution. Dissolve the powder thoroughly by a gentle mix (it can be removed by centrifuge if any precipitate in the solution is found). Pipette 300 µl 1x Assay Diluent into each tube. Use the Positive Control (P-1) solution to produce a dilution series (shown below). Mix each tube thoroughly before the next transfer. 1x Assay Diluent serves as the background. (See i. Positive Control of part IX. TYPICAL DATA for a typical result on page 9).



4. If the Wash Concentrate (20x) (Item B) contains visible crystals, warm to room temperature and mix gently until dissolved. Dilute 20 ml of Wash Buffer Concentrate into deionized or distilled water to yield 400 ml of 1x Wash Buffer.

5. Briefly spin the detection antibody (Item C-1 or C-2) before use. Add 100 µl of 1x Assay Diluent into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4°C for 5 days or at -80°C for one month). The rabbit anti-mTOR (S2448) or biotinylated anti-mTOR antibody concentrate should be diluted 55-fold with 1x Assay Diluent and used in step 4 of Part VII Assay Procedure.

6. Briefly spin the HRP-conjugated anti-rabbit IgG (Item D-1) or Streptavidin (Item G) before use. HRP-conjugated anti-rabbit IgG concentrate should be diluted 1000-fold with 1x Assay Diluent. HRP-conjugated Streptavidin concentrate should be diluted 300-fold with 1x Assay Diluent.

   For example: Briefly spin the vial. Add 5 µl of HRP- conjugated anti-rabbit IgG concentrate into a tube with 5.0 mL 1x Assay Diluent, pipette up and down to mix gently to prepare a 1000-fold diluted HRP-conjugated anti-rabbit IgG solution. Mix well.

7. Cell Lysate Buffer should be diluted 2-fold with deionized or distilled water before use (recommend to add protease and phosphatase inhibitors).

VII. ASSAY PROCEDURE:

1. Bring all reagents to room temperature (18 - 25°C) before use. It is recommended that all samples or Positive Control should be run at least in duplicate.

2. Add 100 µl of each sample or positive control into appropriate wells. Cover well with plate holder and incubate for 2.5 hours at room temperature or over night at 4°C with shaking.

3. Discard the solution and wash 4 times with 1x Wash Solution. Wash by filling each well with Wash Buffer (300 µl) using a multi-channel pipette or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.

4. Add 100 µl of prepared 1x detection antibody, anti- mTOR (S2448) or anti-mTOR (Reagent Preparation step 5) to appropriate wells. Incubate for 1 hour at room temperature with shaking.

5. Discard the solution. Repeat the wash as in step 3.

6. Add 100 µl of prepared 1x HRP-conjugated anti-rabbit IgG against anti- mTOR (S2448) or Streptavidin against anti-mTOR (see Reagent Preparation step 6) to corresponding wells. Incubate for 1 hour at room temperature with shaking.

7. Discard the solution. Repeat the wash as in step 3.

8. Add 100 µl of TMB One-Step Substrate Reagent (Item H) to each well. Incubate for 30 minutes at room temperature in the dark with shaking.

9. Add 50 µl of Stop Solution (Item I) to each well. Read at 450 nm immediately.

VIII. ASSAY PROCEDURE SUMMARY

1. Prepare all reagents, samples and standards as instructed.

⇩

2. Add 100 µl sample or positive control to each well.
   Incubate 2.5 hours at room temperature or overnight at 4°C.

⇩

3. Add 100 µl prepared primary antibody to each well.
   Incubate 1.0 hours at room temperature.

⇩

4. Add 100 µl prepared 1X HRP-Conjugated antibody solution.
   Incubate 1 hour at room temperature.

⇩

5. Add 100 µl TMB One-Step Substrate Reagent to each well.
   Incubate 30 minutes at room temperature.

⇩

6. Add 50 µl Stop Solution to each well.
   Read at 450 nm immediately.

IX. TYPICAL DATA

ELISA data analysis: Average the duplicate readings for each sample or positive.

i. Positive Control

A431 cells were treated with recombinant human EGF at 37^{o}C for 20 min. Solubilize cells at 4 x 10^{7} cells/ml in Cell Lysate Buffer. Serial dilutions of lysates were analyzed in this ELISA. Please see step 3 of Part VI Reagent Preparation for detail.



ii. Recombinant Human EGF Stimulation of A431 Cell Lines

A431 cells were treated or untreated with 100 ng/ml recombinant human EGF for 20 min. Cell lysates were analyzed using this phosphoELISA and Western Blot.

A). ELISA



B). Western-Blot Analysis



X. TROUBLESHOOTING GUIDE

Problem	Cause	Solution
1. Sample signals:		
a. Too low	a. Sample concentration is too low	a. Increasing sample concentration
b. Too high	b. Sample concentration is too high	b. Reducing sample concentration
2. Large CV	a. Inaccurate pipetting	a. Check pipettes
3. High background	a. Plate is insufficiently washed	a. Review the manual for proper washing. If using an automated plate washer, check that all ports are unobstructed.
	b. Contaminated wash buffer	b. Make fresh wash buffer
4. Low positive control signal	a. Improper storage of the ELISA kit	a. Upon receipt, the kit should be stored at -20^{o}C. Store the positive control at -70^{o}C after reconstitution.
	b. Stop solution	b. Stop solution should be added to each well before measurement and read OD immediately.
	c. Improper primary or secondary antibody dilution	c. Ensure correct dilution

RayBio® ELISA kits:

Choose from over 1,000 ELISA kits for human, mouse, rat and a variety of other species. Visit www.raybiotech.com for the complete list.

RayBiotech, Inc., the protein array pioneer company, strives to research and develop new products to meet demands of the biomedical community. RayBio's patent-pending technology allows detection of over 400 cytokines, chemokines and other proteins in a single experiment. Our format is simple, sensitive, reliable and cost effective. Products include: Cytokine Arrays, Chemokine Arrays, ELISA kits, Phosphotyrosine kits, Recombinant Proteins, Antibodies, and custom services.

This product is for research use only.

©2004 RayBiotech, Inc.